

YOUR HOSPITAL REDEVELOPMENT PROJECT EQUIPMENT COMMISSIONING COORDINATION BY ROOM AND ITEM

TO RM NUMBER: **BIOMED** TO RM NAME: BIOMED TO DEPT: BIOMED
 SUPPLIER ITEM #: 90449, 119-0251-00 QTY: 1 PO #: RDVTP-0203MRD07-06 PO LINE #: 1
 ITEM DESCRIPTION: Dual channel bedside printer module. Includes one package z-fold paper. Not compatible with 90481 EEG module. (Includes PO line item #2, power supply).
 VENDOR NAME: Datex-Ohmeda Canada Inc. ADDRESS: #2, 1093 Meyerside Drive, Mississauga, ON L5T 1J6
 VENDOR CONTACT: Peter Robertson VENDOR TEL: (800) 268-1472 VENDOR FAX: (905) 565-8592

ACTIVITY	DATE PLANNED	ASSIGNED TO	DATE COMPLETED	COMPLETED BY	SIGNATURE	REMARKS
DELIVERY	07/02/02	MM				Final Room Location
INSTALLATION		Vendor				by vendor
TRAINING		Vendor				for Operation; for Maintenance
ACCEPTANCE		Biomed				

Acceptance Remarks: Min 1 yr warranty of parts & labour; Manuals for Operator; Manuals for Maintenance; Acceptance Testing by Biomed; Biomed to refer to PO for detailed breakdown of accessories.

GREEN PROJECT TAG: **ASSET TAG:** **SERIAL #:**

Copy of this item with additional inventory tag numbers attached (please indicate in box):

DEFICIENCIES: